

Calibrating Electronic Testers, how to

Drs. Foster & Smith Educational Staff

fast, accurate, and dependable

TEST RESULTS



Calibration establishes standardized reference points to help maintain accurate readings when using [electronic testers](#). Calibration is crucial for accuracy, so it is important to calibrate the instrument regularly, especially if a high level of accuracy is required.

preparation

Pour a small amount of the calibration solution into a clean beaker. If possible, use plastic or glass beakers to minimize any electromagnetic interferences. For accurate calibration and to minimize cross-contamination, use two beakers. One should contain [Rinse Solution](#) for rinsing the electrode or probe and the other should contain the [calibration solution](#).

procedure

1. Immerse electrode or probe into the calibration solution of your choice and stir gently.
2. Conduct reading, per manufacturer's instructions.
3. After reading is stable, confirm/set the calibration.
4. Rinse electrode and repeat process to establish the parameters for the electronic tester (low range and the high range).

5. Your electronic tester is now calibrated and ready for use.

The calibration of electronic testers will vary slightly from product to product but will rely on the basic principles explained here.

Essentials Make it Simple



[Phosphate
Photometers](#)



[Calibration
Solutions](#)



[Hanna Instruments
Testers](#)



[Pinpoint
pH Monitor](#)